

Information Data Sheet

Detector Tube **Cl₂-0.2**
Part No. (US): **803944**
Part No. **D5085801**

1. Application

Detection of chlorine (Cl₂) in air or in technical gases.

2. Detector Tube Sampling Pump

KWIK DRAW (formerly MSA AUER) Gas-Tester® IIH, Gas-Tester® I / ThumbPump™-Sampler, Toximeter® II, Uniphos Kwik-Draw™-Pump or other suitable detector tubes pumps. Observe respective instructions for use.

3. Measuring Range

0.2 ppm ... 3 ppm chlorine at n=10 (10 strokes)


Any number of strokes from 1 to 10 can be chosen. For calculation use n=10 scale, relate as follows:

$$\text{ppm Cl}_2 = \frac{10 \times \text{indication on n=10 scale}}{\text{no. of strokes chosen}}$$

4. Chemical Reaction and Color Change

Oxidation of an aromatic amine by chlorine.

Color change: white → orange / yellow.



5. Ambient Conditions During Sampling

Detector tubes can be used without compensation of reading between 0 °C and 40 °C (32 °F and 104 °F) and between 10 % rh (0.5 g/m³ at 0 °C [32 °F]) and 80 % rh (41 g/m³ at 40 °C [104 °F]).

6. Interferences and Cross Sensitivities

- No interference from:
 - methane, ethane, propane, carbon monoxide.
 - hydrogen, carbon dioxide up to 10 vol. % (n=10).
 - olefinic hydrocarbons (e.g. ethylene) aromatic hydrocarbons (e.g. benzene), hydrogen chloride, phosgene up to 1000 ppm (n=10).
- Sulfur dioxide, hydrogen sulfide, hydrogen cyanide, ammonia are not indicated but will decrease stain length of chlorine indication even when concentrations correspond to measuring range of detector tube.
- Chlorine dioxide, bromine, ozone, nitrogen dioxide will be indicated. The sensitivity of indication is different.
Up to 3 ppm (n=10) the detector tube scale is applicable for bromine as well as for chlorine. Higher concentrations of bromine are indicated.

Manufactured By:

Uniphos Envirotronic Pvt. Ltd (100% EOU)

P.O. Nahuli - 396 108

Tal. Umbergaon,

Dist: Valsad, Gujarat, India

Tel. : +91(260) 2730156 / 57 / 58

Fax : +91(260) 2730160

gasdetection@uniphos-envirotronic.com

Marketing Office:

Uniphos Envirotronic Pvt. Ltd

Readymoney Terrace, 167,

Dr. Annie Besant Road, Worli,

Mumbai 400 018, India.

Tel. : +91(22) 6123 3500

Fax : +91(22) 2493 8826/+91 (22) 2497 8119

www.kwikdrawtubes.com