

Information Data Sheet

Detector Tube Formaldehyde-0.1

Part No. (US): D7086862

Part No. D7086862

1. Application

Detection of formaldehyde (HCHO) in air or in technical gases.

2. Detector Tube Sampling Pump

KWIK DRAW Gas-Tester® I/H, Gas-Tester® I / ThumbPump™-Sampler,
Observe respective instructions for use.

3. Measuring Range

0.2 ppm to 2.5 ppm formaldehyde at n=10 (10 strokes).

0.1 ppm to 1.25 ppm formaldehyde at n=20 (20 strokes).

4. Chemical Reaction and Color Change

Reaction of formaldehyde with xylene and sulfuric acid form a quinoid reaction product.

Color change: White → Violet red



5. Ambient Conditions During Sampling

Detector tubes can be used without compensation of the reading between 5°C and 30°C (41°F and 86°F) and up to 70% RH at 30s°C (86°F)

6. Interferences and Cross Sensitivities

Styrene, Acetaldehyde, Acrolein, Diesel and α -pinene are indicated by yellow brown discoloration, however, with different sensitivity. The lower detection limit for α -pinene is 0.5 ppm for n=20.

No interference of the readings with 500 ppm n-Octane.

Manufactured By:

Uniphos Envirotronic Pvt. Ltd (100% EOU)

P.O. Nahuli - 396 108

Tal. Umbergaon,

Dist: Valsad, Gujarat, India

Tel. : +91(260) 2730156 / 57 / 58

Fax : +91(260) 2730160

gasdetection@uniphos-envirotronic.com

Marketing Office:

Uniphos Envirotronic Pvt. Ltd

Readymoney Terrace, 167,

Dr. Annie Besant Road, Worli,

Mumbai 400 018, India.

Tel. : +91(22) 6123 3500

Fax : +91(22) 2493 8826/+91 (22) 2497 8119

www.kwikdrawtubes.com