

Uniphos Detector Tube Data Sheet

Direct Reading Length of Stain Type

t-Butyl Mercaptan C₄H₉SH

TLV (TWA): 0.5 ppm

TLV (STEL): ----

SBM-1M

Flammable Range: ----

Specification:

Measurement Range	Extended	Standard	Extended
	Do not extend	0.5 - 15 mg/m ³	Do not extend
No. of Pump Strokes	3 (300 mL)		
Volume Correction Factor (VCF)*	1.0		
Sampling Time	1.5 minutes per pump stroke (100 mL)		
Color Change	Yellow → Pink		
Shelf Life	1 year		
Storage condition	Below 10°C (50°F)		
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.		
Reaction Principle	C ₄ H ₉ SH + HgCl ₂ → CH ₃ S.HgCl + HCl HCl + Base → Chloride		
Relative standard deviation	± 10 – 15 %		

* Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

Possible Interferences:

Compound	Concentration (mg/m ³)	Interference	Colour change / Comments
Other Mercaptans		+	Pink
Phosphine		+	Pink
Hydrogen sulphide		+	Pink
Ammonia	15	-	No effect
Sulphur dioxide	30	-	No effect
Carbon dioxide	100	-	No effect

Other measurable compounds:

Compound	Correction Factor	No. of pump strokes	Measuring range
Ethyl Mercaptan	1	3	0.5 - 15 mg/m ³
Methyl Mercaptan	1	3	0.5 - 15 mg/m ³

Correction for Environmental Parameters:

Temperature	Not necessary between 0 - 40°C (32 - 104°F).
Relative humidity	Not necessary between 10 - 90 %.



Manufactured By: Uniphos Envirotronic Pvt. Ltd.
Nahuli (PO), Umbergaon (Tal.)
Valsad (Dist), Gujarat, India – 396108

Rev.0920