

**Specification :**

	Extended	Standard	Extended
Measurement Range	Do not extend	0.02 – 0.3 %	Do not extend
No. of Pump Strokes		1 (100 ml)	
Volume Correction Factor (VCF)*		1.0	
Sampling Time	1 minute per pump stroke (100 ml)		
Color Change	Yellow → Pale blue		
Detection Limit	0.01%		
Shelf Life	3 years		
Relative standard deviation	± 10 – 15 %		

* Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

Reaction Principle :**Possible Interferences :**

Compound	Concentration (ppm)	Interference	Color Change / Comments
Hydrogen sulphide	≥ 100	+	Blue ring
Carbon monoxide	≥ 2000	+	Light blue
Carbon dioxide	3 %	No	No Effect

Correction For Environmental Parameters :

Temperature	Not necessary between 0 - 40°C (32 - 104°F).
Relative humidity	Not necessary between 10 - 90 %.

Calibration Of The Tube :

Static gas dilution method.