

# Uniphos Detector Tube Data Sheet

Direct Reading Length of Stain Type

## Acetylene

## C<sub>2</sub>H<sub>2</sub>

## SAL-5

TLV (TWA): -----

TLV (STEL):-----

Flammable Range: 2.5 – 81 %

### Specification:

Measurement Range	Extended	Standard	Extended
	Do not extend	100 - 5000 ppm	Do not extend
No. of Pump Strokes		1 (100 mL)	
Volume Correction Factor (VCF)*		1.0	
Sampling Time	1 minute per pump stroke (100 mL)		
Color Change	White → Brown with green top		
Shelf Life	1 year		
Storage condition	Below 10°C (50°F)		
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.		
Reaction Principle	C <sub>2</sub> H <sub>2</sub> + I <sub>2</sub> O <sub>5</sub> + H <sub>2</sub> SO <sub>7</sub> → I <sub>2</sub>		
Relative standard deviation	± 10 to 15 %		

\* Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

### Possible Interferences:

Compound	Concentration (ppm)	Interference	Colour change / Comments
Benzene		+	Stain movement
1,1,1 Trichloroethane	500	No	No Effect
Phosphine	120	No	No Effect
Ethylene		+	Stain movement

### Correction for Environmental Parameters:

Temperature	Not necessary between 0 - 40°C (32 - 104°F).
Relative humidity	Not necessary between 10 - 90 %.



**Manufactured By:** Uniphos Envirotronic Pvt. Ltd.  
Nahuli (PO), Umbergaon (Tal.)  
Valsad (Dist), Gujarat, India – 396108

Rev.0620