

# Uniphos Detector Tube Data Sheet

Direct Reading Length of Stain Type

## Ammonia

## NH<sub>3</sub>

## SAM-3

TLV (TWA): 25 ppm

TLV(STEL): 35 ppm

Flammable Range: 15-28%

### Specification:

Measurement Range	Extended	Standard	Extended
	2.5 - 5 ppm	5 - 100 ppm	100 - 200 ppm
No. of Pump Strokes	2 (200 mL)	1 (100 mL)	0.5 (50 mL)
Volume Correction Factor (VCF)*	0.5	1.0	2.0
Sampling Time	1 minute per pump stroke (100 mL)		
Color Change	Yellow → Blue		
Shelf Life	2 years		
Storage condition	Below 10°C (50°F)		
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.		
Reaction Principle	NH <sub>3</sub> + Acid → Reaction product		
Relative standard deviation	± 10 to 15 %		

\* Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

### Possible Interferences:

Compound	Concentration (ppm)	Interference	Colour change / Comments
Amines		+	Blue
Hydrazine		+	Blue
Carbon dioxide	50000	No	No Effect

### Other Measurable Compounds:

Compound	Gas Correction factor (GCF)*	No. of pump strokes	Measuring range
n-Butyl amine	1	1	5 – 100 ppm
Diethyl amine	0.7	1	3.5 – 70 ppm

\* Multiply the observed reading by the correction factor (VCF x GCF) to obtain the true concentration.

### Correction for Environmental Parameters:

Temperature	Not necessary between 0 - 40°C (32 - 104°F).
Relative humidity	Not necessary between 10 - 90 %.



Manufactured By: Uniphos Envirotronic Pvt. Ltd.  
Nahuli (PO), Umbergaon (Tal.)  
Valsad (Dist), Gujarat, India – 396108

Rev.0620