

Uniphos Detector Tube Data Sheet

Direct Reading Length of Stain Type

Carbon Dioxide

CO₂

SCD-8

TLV (TWA): 5000 ppm

TLV (STEL): 30,000 ppm

Non-Flammable

Specification:

Measurement Range	Extended	Standard	Extended
	0.5 - 1%	1 - 20%	20 - 40%
No. of Pump Strokes	2 (200 mL)	1 (100 mL)	0.5 (50 mL)
Volume Correction Factor (VCF)*	0.5	1.0	2.0
Sampling Time	2.5 minutes per pump stroke (100 mL)		
Color Change	Blue → Off white		
Shelf Life	2 years		
Storage condition	Below 10°C (50°F)		
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.		
Reaction Principle	CO ₂ + NaOH → Na ₂ CO ₃ + H ₂ O		
Relative standard deviation	± 10 – 15 %		

* Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

Possible interferences:

Compound	Concentration (%)	Interference	Colour change / Comments
Carbon Monoxide	1.5	+	Color darkens
Hydrogen Sulphide	0.2	+	
Sulphur dioxide	0.02	+	

Correction for Environmental Parameters:

Temperature (°C):	4	20	30	40
Temperature Correction Factor (TCF) *	0.93	1	1.06	1.24

Relative humidity	Not necessary between 10 - 90 %.
-------------------	----------------------------------

* Multiply the observed reading by the correction factor (VCFXTCF) to obtain the true concentration.



Manufactured By: Uniphos Envirotronic Pvt. Ltd.
Nahuli (PO), Umbergaon (Tal.)
Valsad (Dist), Gujarat, India – 396108

Rev.0720