

# Uniphos Detector Tube Data Sheet

Direct Reading Length of Stain Type

## Chlorine

TLV (TWA): 0.5 ppm

## Cl<sub>2</sub>

TLV (STEL): 1 ppm

## SCL-3

Flammable Range: NA

### Specification:

Measurement Range	Extended	Standard	Extended
	2.5 - 5 ppm	5 - 100 ppm	100 - 200 ppm
No. of Pump Strokes	2 (200 mL)	1 (100 mL)	0.5 (50 mL)
Volume Correction Factor (VCF)*	0.5	1.0	2.0
Sampling Time	1.5 minutes per pump stroke (100 mL)		
Color Change	White → Orange		
Shelf Life	2 years		
Storage condition	Below 10°C (50°F)		
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.		
Reaction Principle	Cl <sub>2</sub> + o - Tolidine → Orange		
Relative standard deviation	± 10 – 15 %		

\* Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

### Possible Interferences:

Compound	Concentration (ppm)	Interference	Colour change / Comments
Chlorine dioxide		+	Orange
Nitrogen oxide		+	Diffuse
Nitrogen dioxide		+	Orange
Bromine		+	Orange
Iodine		+	Orange

### Correction for Environmental Parameters:

Temperature	Not necessary between 0 - 40°C (32 - 104°F).
Relative humidity	Not necessary between 10 - 90 %.

**Manufactured By:** Uniphos Envirotronic Pvt. Ltd.  
Nahuli (PO), Umbergaon (Tal.)  
Valsad (Dist), Gujarat, India – 396108



Rev.0720