

Uniphos Detector Tube Data Sheet

Direct Reading Length of Stain Type

1- Butanol

C_4H_9OH

SNB-4L

TLV (TWA): 20 ppm

TLV (STEL):-----

Flammable Range: 1.4-11.2%

Specification:

Measurement Range	Extended	Standard	Extended
	Do not extend	10-150 ppm	Do not extend
No. of Pump Strokes		2 (200 mL)	
Volume Correction Factor (VCF)*		1.0	
Sampling Time	2.5 minutes per pump stroke (100 mL)		
Color Change	Yellow → Pale green		
Shelf Life	1 Year		
Storage condition	Below 10°C (50°F)		
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.		
Reaction Principle	$C_4H_9OH + Cr^{6+} + H_3PO_4 \rightarrow Cr^{3+}$		
Relative standard deviation	± 10 – 15 %		

* Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

Possible Interferences:

Compound	Concentration (ppm)	Interference	Colour change / Comments
Alcohol		+	Pale green

Correction for Environmental Parameters:

Temperature	Not necessary between 0 - 40°C (32 - 104°F).
Relative humidity	Not necessary between 10 - 90 %.



Manufactured By: Uniphos Envirotronic Pvt. Ltd.
Nahuli (PO), Umbergaon (Tal.)
Valsad (Dist), Gujarat, India – 396108

Rev.0720