## **Uniphos Detector Tube Data Sheet**

Direct Reading Length of Stain Type

# **Ethyl benzene**

C<sub>8</sub>H<sub>10</sub>

SEB-4M

TLV (STEL): ----TLV (TWA): 20 ppm Flammable Range: 1-6.7 %

### **Specification:**

Measurement Range	Extended	Standard	Extended
	Do not extend	30 – 400 ppm	Do not extend
No. of Pump Strokes		1 (100 mL)	
Volume Correction Factor (VCF)*		1.0	
Sampling Time	1.5 minutes per pump stroke (100 mL)		
Color Change	White → Brown		
Shelf Life	2 years		
Storage condition	Below 10°C (50°F)		
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.		
Reaction Principle	$C_8H_{10} + I_2O_5 + H_2SO_4 \rightarrow I_2$		
Relative standard deviation	± 10 – 15 %		

<sup>\*</sup> Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

### Possible Interferences:

Compound	Concentration (ppm)	Interference	Colour change / Comments
Benzene		+	Brown
Toluene		+	Brown
Xylene		+	Brown
n - Hexane	300	No	No Effect

### **Correction for Environmental Parameters:**

Temperature	Not necessary between 0 - 40°C (32 - 104°F).
Relative humidity	Not necessary between 10 - 90 %.