

# Uniphos Detector Tube Data Sheet

Direct Reading Length of Stain Type

## Nitrogen Dioxide

NO<sub>2</sub>

SND-2

TLV (TWA): 3 ppm

TLV (STEL): 5 ppm

Flammable Range: NA

### Specification:

Measurement Range	Extended	Standard	Extended
	Do not extend	0.5 - 30 ppm	Do not extend
No. of Pump Strokes		2 (200 mL)	
Volume Correction Factor (VCF)*		1.0	
Sampling Time	2.0 minute per pump stroke (100 mL)		
Color Change	White → Orange		
Shelf Life	2 year		
Storage condition	Below 10°C (50°F)		
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.		
Reaction Principle	NO <sub>2</sub> + o - Tolidine → Orange reaction product		
Relative standard deviation	± 10 – 15 %		

\* Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

### Possible Interferences:

Compound	Concentration (ppm)	Interference	Colour change / Comments
Chlorine		+	Orange
Bromine		+	Orange
Carbon monoxide	3000	No	No Effect
Sulphur dioxide	200	No	No Effect

### Correction for Environmental Parameters:

Temperature	Not necessary between 0 - 40°C (32 - 104°F).
Relative humidity	Not necessary between 10 - 90 %.



**Manufactured By:** Uniphos Envirotronic Pvt. Ltd.  
Nahuli (PO), Umbergaon (Tal.)  
Valsad (Dist), Gujarat, India – 396108

Rev.0621