

# Uniphos Detector Tube Data Sheet

Direct Reading Length of Stain Type

## Nitrogen Oxides

NO<sub>x</sub>

SNO-5M

TLV (TWA ): 3 ppm (NO<sub>2</sub>) & 25 ppm (NO)

Flammable Range: NA

### Specification:

Measurement Range	Extended	Standard	Extended
	Do not extend	100 - 2500 ppm	Do not extend
No. of Pump Strokes		1 (100 mL)	
Volume Correction Factor (VCF)*		1.0	
Sampling Time	1.5 minutes per pump stroke (100 mL)		
Color Change	White → Yellowish orange		
Shelf Life	1 year		
Storage condition	Below 10°C (50°F)		
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.		
Reaction Principle	NO + Cr <sup>6+</sup> + H <sub>2</sub> SO <sub>4</sub> → NO <sub>2</sub> NO <sub>2</sub> + o - Tolidine → Yellowish orange reaction product		
Relative standard deviation	± 10 – 15 %		

\* Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

### Possible Interferences:

Compound	Concentration (ppm)	Interference	Colour change / Comments
Chlorine		+	Yellowish orange
Bromine		+	Yellowish orange
Hydrogen Sulphide	100	No	No Effect
Sulphur dioxide	200	No	No Effect

### Correction for Environmental Parameters:

Temperature	Not necessary between 0 - 40°C (32 - 104°F).
Relative humidity	Not necessary between 10 - 90 %.



**Manufactured By:** Uniphos Envirotronic Pvt. Ltd.  
Nahuli (PO), Umbergaon (Tal.)  
Valsad (Dist), Gujarat, India – 396108

Rev.0920