

# Uniphos Detector Tube Data Sheet

Direct Reading Length of Stain Type

## Water Vapor

TLV (TWA): -----

## H<sub>2</sub>O

TLV (STEL): ----

## SWA-5

Non Flammable

### Specification:

Measurement Range	Extended	Standard	Extended
	Do not extend	0.1– 2 mg/l	Do not extend
No. of Pump Strokes		1 (100 mL)	
Volume Correction Factor (VCF)*		1.0	
Sampling Time	1 minute per pump stroke (100 mL)		
Color Change	Yellow → Greenish blue		
Shelf Life	1 year		
Storage condition	Below 10°C (50°F)		
Calibration condition	Calibrated at 20°C (68°F).		
Reaction Principle	$H_2O + Mg(ClO_4)_2 \rightarrow Mg(ClO_4)_2 \cdot H_2O$		
Relative standard deviation	± 10 – 15 %		

\* Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

### Possible Interferences:

Compound	Concentration (ppm)	Interference	Colour change / Comments
Methane	99.9%	No	No effect
Hydrogen sulphide	2000	No	No effect
Methanol	50	No	No effect
Hexane	3000	No	No effect

### Correction for Environmental Parameters:

Temperature	Not necessary between 0 - 40°C (32 - 104°F).
-------------	--



**Manufactured By:** Uniphos Envirotronic Pvt. Ltd.  
Nahuli (PO), Umbergaon (Tal.)  
Valsad (Dist), Gujarat, India – 396108

Rev.0920