
**Specification :**

Measurement Range	Extended	Standard	Extended
	0.125 – 0.25%	0.25 – 3%	3 – 6%
No. of Pump Strokes	2 (200 ml)	1 (100 ml)	0.5 (50 ml)
Volume Correction Factor (VCF)*	0.5	1.0	2.0
Sampling Time	2.5 minutes per pump stroke (100 ml)		
Color Change	Blue → Off White		
Detection Limit	0.05% (2 pump strokes)		
Shelf Life	3 years		
Relative standard deviation	± 10 – 15 %		

* Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

Reaction Principle :**Possible Interferences :**

Compound	Concentration (ppm)	Interference	Color Change / Comments
Carbon monoxide	≤1.5	No	No Effect
Hydrogen Sulphide	0.2	+	Off White
Ammonia	≤5	No	No Effect
Sulphur dioxide	0.02	+	Off White

Correction For Environmental Parameters :

Temperature (°C):	4	20	30	40
Temperature Correction Factor (TCF) *:	0.82	1	1.03	1.13

Relative humidity	Not necessary between 10 – 90 %.			
-------------------	----------------------------------	--	--	--

* Multiply the observed reading by the correction factor (VCF X TCF) to obtain the true concentration.

Calibration Of The Tube :

Static gas dilution method.