

**Specification :**

	<b>Extended</b>	<b>Standard</b>	<b>Extended</b>
<b>Measurement Range</b>	2.5 – 5 ppm	5 – 100 ppm	100 – 200 ppm
<b>No. of Pump Strokes</b>	2 (200 ml)	1 (100 ml)	0.5 (50 ml)
<b>Volume Correction Factor (VCF)*</b>	0.5	1.0	2.0
<b>Sampling Time</b>	1.5 minute per pump stroke (100 ml)		
<b>Color Change</b>	White → Yellow		
<b>Detection Limit</b>	1 ppm (2 pump strokes)		
<b>Shelf Life</b>	1 year		
<b>Relative standard deviation</b>	± 10 – 15 %		

\* Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

**Reaction Principle :****Possible Interferences :**

<b>Compound</b>	<b>Concentration (ppm)</b>	<b>Interference</b>	<b>Color Change / Comments</b>
1,2 dichloroethylene		+	Yellow
Chlorine		+	Yellow
Hexane	≤ 100	No	No Effect

**Correction For Environmental Parameters :**

<b>Temperature</b>	Not necessary between 0 - 40°C (32 - 104°F).
<b>Relative humidity</b>	Not necessary between 10 - 90 %.

**Calibration Of The Tube :**

Static gas dilution method.